



Instruction Booklet

# pursuit



Serial Number



Code: PC01010004

# Table of Content

Introduction	2
Important Precautions	3
Safety Warning & Instruction Labels	4
Identification of Parts	5
Installation Procedure	8
Charging the Batteries	13
Regular Maintenance and Care	14
Safety Information on Electromagnetic Interference (EMI)	14
Warranty	16
Trouble Shooting	17
Care and Maintenance	18
Specifications	19

Thank you and congratulations on purchasing your new Abilize Mobility Powerchair.

It is designed to provide you with transportation ability indoors and outdoors.

We pride ourselves on providing safe and comfortable products. Our goal is to ensure your complete satisfaction. We sincerely hope you enjoy your Abilize Mobility Powerchair.

Please read and observe all warnings and instructions provided in owner's manual before you operate the various functions of this Powerchair. Also, please retain this booklet for future reference.

## IF YOU HAVE ANY QUESTIONS, YOU CAN CONTACT:

### UK Authorised Dealer

CareCo (UK) Limited,  
1 Turing Court,  
Great Notley,  
Braintree,  
CM77 7AT

0333 015 5000

[www.CareCo.co.uk](http://www.CareCo.co.uk)

### Information of European Representation

**MedNet EC-REP GmbH**  
Borkstrasse 10,  
48163 Muenster,  
Germany

**UKRP** MediMap Ltd, 2 The Drift,  
Thurston, Suffolk, IP31 3RT



CareCo (UK) Limited,  
1 Turing Court, Great Notley,  
Braintree, CM77 7AT



## IMPORTANT PRECAUTIONS

Safety is the main consideration when using your chair. It is required that you read and understand all the operating and safety instructions discussed in this manual, and ensure your chair is correctly fitted and adjusted by your dealer or the prescribing healthcare professional.

Make sure to engage the wheel locks before entering or leaving the chair. The wheel locks are designed to prevent movement of the chair. It is preferable to ensure the front castors are in the forward position before transferring into or out of the chair. With the castors in the forward position the wheel base of the chair is increased and therefore, offers more stability.

DO NOT move forward in the seat while leaning forward out of the chair. If you need to pick an object up from the floor, drive past it, then reverse so the front castors are in the forward position. This gives the chair its greater stability.

To maintain lateral stability do not reach further than the length of your arm. DO NOT lean out of the chair as this will cause instability.

When transferring, DO NOT stand on the leg rests. Depending on the style of leg rests either swing them away or fold them up before transferring.

When approaching a ramp, be sure of your own ability and your limitations in terms of strength and endurance.

### **Before attempting a ramp the following basic rules should be considered:**

1. Surface of the ramp: Is it slippery?
2. Degree of incline: Is it too steep to attempt alone?
3. Length of ramp: Is it too long for your endurance?
4. Obstacles: Are there any obstacles on the ramp that would necessitate assistance?

### **Be very careful when going up or down steep inclines:**

If it becomes necessary to stop when going up an incline, special care must be taken to avoid abrupt or sudden forward movement. During continuous forward movements, the chair is capable of falling backwards.


Always keep the chair under control when going down a ramp or incline. Speed should be controlled at all times.

If you encounter a kerb whilst driving, caution should be taken to prevent the user being thrown forward:

**CAUTION: Do not try to climb a kerb by driving up it. Have a helper move the chair for you following the instructions below.**


Lifting the rear wheels first if going down a kerb, or the front wheels first if going up, ensure both sides are lifted equally, and ensure the motors are disengaged.

**CAUTION: Please be aware that any adjustments on the power chair may affect the handling and performance.**




<p>1</p>	<p><b>Warning Sticker</b></p> <ol style="list-style-type: none"> <li>1. Please read the instruction booklet carefully before using your scooter.</li> <li>2. Do not drive the scooter on slippery surfaces or on slopes over 8 degrees limit.</li> <li>3. Do not drive on highway, crowded roads, or unfamiliar areas.</li> <li>4. Do not turn at high speed in either forward or reverse.</li> <li>5. Do no wash with water or leave scooter in humid environment since water can damage the electronic parts.</li> </ol>
<p>2</p>	<p>REF) Model Number MD) Medical Device SN) Serial Number Date of Manufacture Manufacturer Importer</p> <p>EU Authorized representative WEEE Conformity UKRP Consult Instructions for Use UK Conformity Assessed Warning</p> <p>CE European Conformity Max: Maximum weight capacity</p>
<p>3</p>	<p>Wiring diagram label</p>
<p>4</p>	<p>Transaxle test ok label</p>
<p>5</p>	<p>Left N-D Lever adjustment label which instructs freewheel mode operation.</p>
<p>6</p>	<p>Right N-D Lever adjustment label which instructs freewheel mode operation.</p>

Before attempting to drive this power chair on your own, it is important that you familiarize yourself with the controls and how they operate.



**Figure 1 - Front View**



**Figure 2 - Control Panel / Joystick**




**Figure 3 - Back View**



## IDENTIFICATION OF PARTS

### Footrest Height Adjustment:



1. Remove both side's height fixation screws (A). (See Fig. 4)
2. Adjust to required position.
3. Aim for the required hole's position to tighten height fixation screw.



**CAUTION: Careful not to interfere with front wheel when adjusting.**


### Seat Belt Usage:

1. Prior to driving fasten the seat belt as the picture illustrates (B) (B1).
2. Seat Belt can be unfastened by pressing down the red button (B2).



### Free Wheel Levers:



1. For your convenience, the chair is equipped with two free wheel levers. These levers allow you to disengage the drive motors and manoeuvre the chair manually if you need to stop or encounter a problem.
2. When travelling, set the two metallic levers (C) to the drive position. (See Fig. 5)
3. The battery box with plug goes in the front position. The battery box with reflector goes in the rear position.



**CAUTION: DO NOT use your chair with the drive disengaged unless you have an attendant! DO NOT disengage the drive motors when your chair is on an incline. The chair could roll down on its own, causing injury!**

### Armrest Height Adjustment

1. Use tool adjust armrest height to the required hole's position. (See Fig. 6. E)
2. Aim for the hole's position to tighten Screw (D).



### Extendible Joystick Adjustment:

The joystick can be adjusted so that it is in a convenient position for use. To adjust the unit forward and backwards use the knob underneath the armrest pad. When loosened this allows the horizontal tube to be adjusted to the required position.

To adjust the unit vertically, there is a hexagon key screw in a hexagon nut attached to the tubular construction supporting the joystick. Using the hexagon key provided loosen the screw, adjust unit to the required height and re-tighten screw to secure.



### Joystick conversions to left hand

1. Remove joystick and horizontal tube from right hand arm by releasing hand knob and hexagon screw (G). (See Fig. 13)
2. Remove the two screws holding the support tube to the joystick and rotate the tube through 180° to the opposite side and secure to the joystick (G1) with the screws.
3. To convert to opposite hands remove support brackets from pads and rotate through 180° and align with opposite screw holes and re-secure with screws.
4. Secure the original right hand armrest to the left hand arm and repeat for the other hand.
5. Re-position joystick unit to left hand side of chair and assemble then secure to left hand arm.
6. Adjust heights and position as previous paragraphs.

### Option: Swing-away Joystick Adjustment:

By pushing or pulling the joystick (G2) to desired position.

- Do not drive while operating the above function.
- Watch out for pinch warning.





## INSTALLATION PROCEDURE

1. First read and fully understand this manual.
2. Mount all the electrical parts of the wheelchair system (motors, park brakes, batteries, Power Module, Remote) on the wheelchair. for the physical dimensions of the LiNX LE System Power Module, Remote and mounting recommendations.
3. Do not connect any cables before all the parts of the electrical system are mounted.
4. Connect the LiNX LE System Power Module to the motors the park brakes and the Remote.
5. Connect the LiNX LE System Power Module to the batteries.

**CAUTION: Do not turn on the wheelchair yet.**

**CAUTION: Do not connect the positive terminal (B+) of the battery to the LiNX LE System Power Module until the wheelchair is completely wired and ready for testing as described in the Testing section.**

6. Lift the wheelchair off the ground and check the installation thoroughly.
7. Program the system to the requirements of a particular wheelchair or user.
8. Test the system for functionality and safety.

### Power up/down:

To switch ON the LiNX LE System, press the Power button. The Power button is the only user input that can activate the system.

If the system is healthy, the Status indicator (through the Power button) will light up green, and the Battery Gauge will display the current battery status.



If there is a fault with the system when powering up, the status indicator will indicate the fault with a series of red flashes (see section 3 Diagnostics). If the fault is one that prevents the system from driving, then the battery gauge will flash continuously.

To switch OFF the system, press the Power button; the system will power down and the Status indicator will switch off.

The Power button is also used to perform an EMERGENCY STOP. See next section.

### Emergency Stop:

If the user needs to stop the wheelchair quickly, the Power button can be pressed to perform an EMERGENCY STOP. The wheelchair will come to a halt quickly; the rate is set by the Emergency Deceleration parameter.

**CAUTION: If this parameter is set too high, the user can lose balance or fall out of the chair.**

### Drive Inhibit Indication:

Drive inhibit mode is indicated by the battery gauge with a right-to-left chase sequence.

The chase sequence starts with the green LED on the right-hand side, and one-by-one, each LED will switch on and then off. When the sequence completes at the left-most red LED, it begins

again at the right-hand side.

The chase sequence continues until the error condition has been cleared.

### **OONAPU:**

OONAPU (“Out Of Neutral At Power Up”) is a safety feature that prevents accidental movement of the wheelchair, either when powering up, or when the wheelchair comes out of an inhibit state.

If the LiNX LE System is turned on (or comes out of an inhibit state) while the joystick is not in the centre position, an OONAPU warning is displayed. During an OONAPU warning, the battery gauge LEDs will flash continually to alert the user, and the chair will not drive. If the joystick is returned to the centre position within five seconds, the warning will clear and the wheelchair will drive normally.

However, if the joystick remains out of neutral for longer than five seconds, an OONAPU error will occur; the error is displayed by the Status indicator flashing red, and the chair will not drive. To clear the error, return the joystick to the neutral position and power the unit OFF and then on again.

### **The Joystick:**

The joystick controls the direction and speed of the wheelchair.

When the joystick is deflected from the centre position, the wheelchair will move in the direction of the joystick movement.




The speed of the wheelchair is proportional to the joystick deflection, so that the further the joystick is moved from the centre position, the faster the wheelchair will travel.

### **Controlling Maximum Speed:**

The speed dial allows the user to limit the maximum speed of the wheelchair (that is the speed when the joystick is fully deflected) to suit their preference and environment.

The dial offers 10 discrete steps between the lowest speed (dial set to the left) and the highest speed (dial set to the right).


As a visual reminder, a speed symbol (shown left) is positioned just below the speed dial to indicate the low and high positions of the speed dial.



**CAUTION: It is the responsibility of the wheelchair manufacturer to inform the wheelchair user about the wheelchair’s stopping distances.**

### **The Horn:**

The Horn button is located above the Power button. Press the Horn button to sound the horn. The horn will sound for as long as the Horn button is pressed.



**The Battery Gauge:**

The battery gauge comprises five different LEDs (1 x RED, 2 x AMBER, 2 x GREEN), situated above the Remote’s Horn button. The number of LEDs lit depends on the status of the battery, as shown below.

The battery gauge LEDs are also used to display charging information. See section 2.2 Battery charging for more details.



**Normal Operation:**

Battery Gauge	Battery Level	Notes
	Fully Charged	This level is set by the Battery Gauge Maximum parameter.
	Consider charging battery	
	Battery needs charging	This level is set by the Batt Gauge Minimum parameter.

**High Voltage Warning:**

A high voltage warning is indicated by all LEDs on, and the green LEDs flashing. This occurs when the battery voltage level has risen above the high voltage warning set-point.



**Low Voltage Warning:**


A low voltage warning is indicated with the left-most LED flashing. This occurs when the battery voltage level has decreased below its low voltage warning set-point. Charge the battery immediately.



**Cut-off Voltage:**

When the battery voltage decreases below the battery cut-off voltage:

- The first (red) LED will flash on the battery gauge.
- The status indicator (under the power button) will display a flash code 2 or 7.
- The horn will sound once every 10 seconds.



**The Status Indicator:**

The status indicator is located underneath the power button. When the LiNX LE System is not powered up, the status indicator is not lit.


When the LiNX LE System is powered up, and there are no faults with the system, the status indicator will be lit green.



If, when powered up, there is a fault with the system, then the status indicator will flash red. The number of flashes will indicate the type of error.

### The XLR Charger Connector:

The XLR charger connector, which is located on the right-hand side of the Remote, is used to connect to either a battery charger or the LiNX Access Key.



**CAUTION:** Make sure that the battery charger that is used with the vehicle has a drive inhibit function that is correctly connected for use with the controller. The maximum voltage on the inhibit pin must not exceed 3V if a battery voltage is to be detected when the battery charger is connected. If you are not sure, ask your dealer or vehicle manufacturer.

The XLR charger connector on the Remote is to be used exclusively for the intended purpose. Warranty will be voided if any unauthorised device is connected to this port.

### The LiNX Communications Bus Connector:

The LiNX Communications Bus connector can be found on the lower front of the Remote (see Figure 28: The Remote: user interface and connectors). The LiNX Communications Bus loom plugs directly into this socket, providing the Remote with both power and communication to the power module.




Fig. 28


### Battery Charging:

Plug the battery charger into the Remote's XLR socket.

The Battery Gauge will indicate the system is being charged by cycling between a left-to-right chase sequence, and then displaying the approximate battery charge state at the end of the chase sequence.

Driving is prevented (inhibited) while the system is being charged.

The LE system does not have to be powered up when charging the battery, however, if it is not powered up, then the battery gauge will not display the charging state/ chase sequence.



**CAUTION:** The maximum charging current for the LiNX LE System is 6A.

The wheelchair manufacturer must specify an appropriate battery charger for the batteries used in the wheelchair. The wheelchair manufacturer must also specify the maximum current of any battery chargers to be used with the controller and warn against using battery chargers of higher current ratings. The battery charger must have over-current protection in the form of a non-resettable fuse. It is the responsibility of the wheelchair manufacturer to manage the risks of battery over-charging and any related gas emissions. To protect the wheelchair wiring from over currents while charging the batteries, chargers must have the ability to reduce their current output when electrically shorted.

### Error Indication:

If, when powered up, there is an error with the system, then the status indicator will flash red. The number of flashes will indicate the type of error.

These are described in the table on the following page.




Flash Code	Error Description
1	Remote / joystick error
2	Network or configuration error
3	Left motor error
4	Right motor error
5	Left park brake error
6	Right park brake error
7	Module error (other than Remote)

For more information about the error, and what to do about it, open the logs within one of the programming & diagnostic tools.

### Free Wheel Operation:

To push the Powerchair by hand or in the event of a fault or battery failure the motors/gear boxes can be released allowing the Power chair to be pushed.

The levers should be switched to the free wheel position on the gear box by rotating them through 90°. Use the manual brakes to park the chair.



**CAUTION:** Always re-engage the free wheeling device after use. Failure to do so may result in injury

### Wheels and Tyres:

Regularly check the condition of front and rear tyres for wear and tear. Contact your supplier for replacements if necessary. Your Powerchair may have split rim wheels which are held together by screws or bolts. When these are removed the wheel rim separates for easy access to carry out puncture repairs or tyre replacements.

### Disassembly and Storage:

1. Your Power chair can be disassembled and folded for storage:
2. Fold the two footrest plates upwards (H1) and press (H) button to swing away.
3. Press the right button (I) to release the support bracket (J)



4. Disconnect plug and socket between battery boxes.
5. Pull off velcro strap (L). Pull to slide battery boxes from sliders.
6. Pull off velcro (M). Using both hands, grip front and rear of the seat upholstery (N) and pull up to fold the chair.
7. To unfold the chair, place each hand on both seat tubes and push downward firmly. Slide battery boxes into position, secure in place by velcro strap and reconnect cables.

### **Batteries and Charging:**

When your batteries are fully charged you should have sufficient power to give you all the mobility required in a day. It is important that you understand how your batteries and charger work.

Batteries must be charged before using the power chair for the first time and are recommended to be charged up to 10 - 14 hours after each day's use.

### **Maintain the Batteries:**

1. Batteries should be charged every night in a well ventilated room.
2. DO NOT place the power wheelchair near radiators or open fireplaces while charging.
3. DO NOT smoke or permit open flames in the immediate vicinity.
4. Turn the chair controller power off before charging.
5. It is advisable that the batteries be charged for a minimum of 10 hours per night to ensure full battery storage capacity. The battery charger is an automatic current limiting device and will shut off when the batteries are fully charged.

### **Charging the Batteries:**

1. Position power chair next to a standard wall outlet.
2. Connect the battery charger to the wheelchair input battery charging socket, which is located on the front of the controller.
3. Connect the battery charger to a standard wall power outlet.
4. Switch the power on.

<b>Indicator Functions:</b>	With bi-coloured LED light Flash-rapid Red = Faulty Flash Red = Standby Solid Red = Charging Solid Green = Fully Charged Yellow = Power
-----------------------------	--

**CAUTION:** Do not use batteries other than the recommended type for your chair and never use a charger other than the one supplied for that purpose. If the chair is not used for a long period of time arrange to have the batteries charged for at least one day (10-14 hours) every month, minimum. Periodically, check that the battery terminals are clean and the connections are tight. Smear a thin film of petroleum jelly on the terminals to guard against corrosion. Always wash your hands after handling batteries.



## **Maintenance and Cleaning:**

An electric wheelchair needs some basic attention to ensure it provides reliable service. We recommend that the user ensures that the power wheelchair is checked regularly for maintenance requirements and receives a thorough, annual maintenance check up.

### **Annual Maintenance:**

We recommend that the chair has at least one full service per year from an authorized dealer. This will ensure that your power chair is functioning properly and also helps prevent future complications. This should include:

1. Checking the tires.
2. Checking the batteries and terminals.
3. Checking the controller program for the user's needs.
4. Checking the frame.
5. Checking the upholstery condition.

### **Regular Maintenance and Cleaning:**

1. Avoid knocking or bumping the controller, especially the joystick.
2. Avoid prolonged exposure of your power chair to extreme conditions, such as heat, cold, or moisture.
3. Keep the controller clean.
4. Check that all controller connectors are tight and secured properly.
5. Never hose off your power chair or place it in direct contact with water.
6. Keep the upholstery and frame clean by wiping with a soft cloth, particularly after driving through wet, sandy or muddy conditions. Do not use harsh abrasive materials when cleaning. Do not apply liquid cleaners or solvents directly to the control box, battery charger or any electrical connections.
7. Keep wheels free from lint, hair, sand and carpet fibres.
8. Lightly oil axle pin, wheel axles and bearings once every three months, if necessary.
9. Inspect the tires. Tread depth should not be allowed to drop below 1/16".
10. Use only recommended batteries and have batteries changed by Qualified Dealer if you have doubts about your ability to lift the components.
11. Charge batteries regularly. Make sure the charger lead plugs are engaged properly in the sockets. Do not disconnect by pulling the cord.
12. With the controller turned off, check the joystick. Make sure it is not bent or damaged and that it returns to the center when you release it. Check the rubber boot around the base of the joystick for damage. Visually inspect the boot. Do not handle or try to repair it. See your authorized dealer for any questions.
13. Visually inspect the controller harnesses. Make sure that they are not frayed, cut or have any wires exposed. See your authorized dealer if there is a problem with any of these harnesses.
14. Ensure that all parts of the controller system are securely fastened to your chair. Do not over tighten any screws.

## **ELECTROMAGNETIC INTERFERENCE AND WARNINGS**

**CAUTION:** It is very important that you read this information regarding the possible effects of Electromagnetic Interference on your power chair.

Powered wheelchairs and motorized scooters may be susceptible to electromagnetic interference (EMI), which is interfering electromagnetic energy (EM) emitted from sources such as radio

stations, TV stations, amateur radio (HAM) transmitters, two-way radios, and cellular phones. The interference (from radio wave sources) can cause the power chair to release its brakes, move by itself, or move in unintended directions. It can also permanently damage the power chair control system. The intensity of the interfering EM energy can be measured in volts per meter (V/m). Each power chair can resist EMI up to a certain intensity. This is called its "immunity level." The higher the immunity level, the greater the protection will be. At this time, current technology is capable of achieving at least a 20 V/m immunity level, which would provide useful protection from the more common sources of radiated EMI. The immunity level of this product is 20 V/m.

There are a number of sources of relatively intense electromagnetic fields in the everyday environment. Some of these sources are obvious and easy to avoid. Others are not apparent and exposure is unavoidable. However, we believe that by following the warnings listed below, your risk to EMI will be minimized.

### **The sources of radiated EMI can be broadly classified into three types:**

1. Hand-held portable transceivers (transmitters-receivers) with the antenna mounted directly on the transmitting unit. Examples include: citizens band (CB) radios, "walkie talkie," security, fire, and police transceivers, mobile telephones, and other personal communication devices.

CAUTION: Some mobile telephones and similar devices transmit signals while they are ON, even when not being used.

2. Medium-range mobile transceivers, such as those used in police cars, fire trucks, ambulances, and taxis. These usually have the antenna mounted on the outside of the vehicle.
3. Long-range transmitters and transceivers such as commercial broadcast transmitters (radio and TV broadcast antenna towers) and amateur (HAM) radios.

CAUTION: Other types of hand-held devices, such as cordless phones, laptop computers, AM/FM radios, TV sets, CD players, and small appliances, such as electric shavers and hair dryers, so far as we know, are not likely to cause EMI problems to your power chair.

### **Power Chair Electromagnetic Interference:**

Because EM energy rapidly becomes more intense as one moves closer to the transmitting antenna (source), the EM fields from hand-held radio wave sources (transceivers) are of special concern. It is possible to unintentionally bring high levels of EM energy very close to the power chair control system while using these devices. This can affect power chair movement and braking. Therefore, the warnings listed below are recommended to prevent possible interference with the control system of the power chair.

**CAUTION: Electromagnetic interference (EMI) from sources such as radio and TV stations, amateur radio (HAM) transmitters, two-way radios, and mobile phones can affect the power chair. Following the warnings listed on the following page should reduce the chance of unintended brake release or power chair movement, which could result in serious injury.**

1. Do not operate hand-held transceivers (transmitters-receivers), such as citizens band (CB) radios, or turn ON personal communication devices, such as mobile phones, while the power chair is turned ON;
2. Be aware of nearby transmitters, such as radio or TV stations, and try to avoid coming close to them;
3. If unintended movement or brake release occurs, turn the power chair OFF as soon as it is safe;
4. Be aware that adding accessories or components, or modifying the power chair, may make it more susceptible to EMI; and
5. Report all incidents of unintended movement or brake release to the distributor listed on the inside front cover of this manual. Note whether there is a source of EMI nearby.

### **Important Information:**






1. 20 volts per meter V/m is a generally achievable and useful immunity level against EMI (as of May 1994). The higher the level, the greater the protection.
2. The immunity level of this product is 20 V/m.

### **Warranty:**

There is a comprehensive twelve-month warranty from the date on which your new scooter is delivered. The warranty covers the scooter for repairs or replacement during this period. For more details, please see the warranty conditions below.

- Any work or replacement part installation must be carried out by an authorised service agent.
- To apply to warranty should your scooter require repair, please contact the authorised dealer.
- Should any part of the scooter require repair or full or part replacement, as a result of a manufacturing or material defect within warranty period, the work will be carried out free of charge. Warranty period:
  1. Frame: 2 year limited warranty.
  2. Electronic parts: 12 months limited warranty.
  3. Batteries: 6 month limited warranty.
- Any repaired or replaced parts will be covered by this warranty for the balance of the warranty period on the powerchair.
- Parts replaced after the original warranty has expired will be covered by a three months warranty.
- Consumable items supplied will not generally be covered during the normal warranty period unless such items require repair or replacement clearly as a direct result of a manufacturing or material defect. Such items include (among others): upholstery, tyres and batteries.
- The above warranty conditions apply to brand new powerchair purchased at the full retail price. If you are unsure whether your powerchair is covered, check with the authorised dealer.
- Under normal circumstances, no responsibility will be accepted where the scooter has failed as a direct result of:
  1. The scooter part not having been maintained in accordance with the manufacturer's recommendations.
  2. Failure to use the manufacturer's specified parts.
  3. The scooter or part having been damaged due to neglect, accident or improper use.
  4. The scooter or part having been altered from the manufacturer's specifications or repairs having been attempted before the service agent is notified.

The manufacturer reserves the right to alter, without notice, any weights, measurements or other technical data shown in this manual. All figures, measurements and capacities shown in this manual are approximate and do not constitute specifications.

Error code	Error Description	Solution
 1	Remote / joystick error	Check cables and connectors Replace Remote
 2	Network or configuration error	Recharge the battery Check charger Check cables and connectors Check Bluetooth pairing Replace modules Contact supplier
 3	Left motor error	Check cables and connectors Replace Power Module Check and/or replace left motor
 4	Right motor error	Check cables and connectors Replace Power Module Check and/or replace right motor
 5	Left park brake error	Check cables and connectors Check left park brake is released Replace Power Module
 6	Right park brake error	Check cables and connectors Check right park brake is released Replace Power Module
 7	Module error (other than Remote)	Check cables and connectors Check modules Replace LiNX Access Key Replace Power Module Recharge battery If chair stalled, reverse away or remove obstacles, or if chair has been moved while turned off, cycle power.

#### Further issues:

If the above troubleshooting does not help to resolve the issue please contact the authorised dealer (page 2). Please have your serial no. to hand to provide to the authorised dealer, which can be found on the battery pack.

## DAILY CHECK

Please always check your scooter before you start your every journey.

Check point	Inspection	What to do if the inspection is failed
N-D lever	Check for correct function	Contact your dealer.
Horn	Check for correct function	Contact your dealer.
Throttles	Pull the wigwag to test the scooter movement	Contact your dealer.
Electro-mechanical brake and Emergency hand brake	Pull the wigwag a little bit and release it to test if brake works. If your scooter comes with emergency handbrake, please check it as well.	Contact your dealer.
Battery Gauge	Check if the battery gauge is displayed and whether it is at low power.	<ul style="list-style-type: none"><li>• Contact your dealer if battery gauge is not working.</li><li>• Recharge the battery immediately if low</li></ul>
Rear mirror (s)	Check if the parts are clean and firmly tighten to the scooter and do not wobble.	<ul style="list-style-type: none"><li>• Clean up the dirt by damp cloth.</li><li>• Tighten the screw or clamping stem that holds the mirror(s)</li></ul>
Lighting	Check if all lights, such as head lights, rear lights, and turn signal are functioning correctly.	Contact your dealer.

## WEEKLY CHECK

Check point	Inspection	What to do if the inspection is failed
Speed Dial Knob	Check for correct function	Contact your dealer.
Armrests	Check if the parts are clean and firmly tighten to the scooter and do not wobble. Tighten the screw knob that holds the armrest.	Contact your dealer.
Wheels/Tires	Inflate the tire to the correct pressure and check that <ol style="list-style-type: none"><li>1. Drive wheels rotate without wobbling.</li><li>2. Tire tread depth is not less than 0.5mm.</li><li>3. No foreign objects in tires.</li></ol>	Contact your dealer.
Motor	Check for correct function	Contact your dealer.
Battery Charger	Check if the charger is functioning correctly and the batteries are charged.	Contact your dealer.

## MONTHLY CHECK

Check point	Inspection	What to do if the inspection is failed
Seat / Upholstery	Check for movement and if it's worn	Contact your dealer.
Electronics	Check if all the battery cables and connectors are firmly tighten to the scooter	Contact your dealer.

Specification	
<b>Overall length</b>	1040 mm / 41"
<b>Overall width</b>	595 mm / 23.4" - 645 mm / 25.4" - 695 mm / 27.4"
<b>Overall height</b>	990 mm / 40"
<b>Front wheels</b>	200 mm / 7.9"
<b>Rear wheels</b>	320 mm / 13"
<b>Weight w/batteries</b>	63.2 kg / 139.3lbs - 63.5 kg / 140lbs - 65 kg / 143.5lbs
<b>Battery Charger</b>	5A Off Board
<b>Max. speed</b>	8 kmph / 5 mph
<b>Weight capacity</b>	136 kg / 300lbs - 136 kg / 300lbs - 145 kg / 320lbs
<b>Ground clearance</b>	80 mm / 3.2"
<b>Grade climbable</b>	8 degree
<b>Curb climbable</b>	25 mm / 1"
<b>Turning radius</b>	850 mm / 33.5"
<b>Brake</b>	Electro-Mechanical
<b>Seat type</b>	Padded Breathable Ballistic Nylon Back Rest W/ Adjustable Back Angle & Seat Belt
<b>Seat Back Angle</b>	100 degree
<b>Seat Back Height</b>	470 mm / 18.5"
<b>Seat width</b>	410 mm / 16" - 460 mm / 18" - 510 mm / 20"
<b>Motor size</b>	420W 4600 r.p.m
<b>Battery size</b>	(2) 12V. 36Ah
<b>Battery weight</b>	25 kg / 55 lbs
<b>Travel range</b>	35 km / 21.7 Miles
<b>Battery charger</b>	19 km / 12 Miles

\*Subject to change without notice. (Issue E 13/01/2023 )









